



Date and Time : 2014-6-10 9:02:30
Sample : Clen Sample
MW :
Lot No. :

Probe: ESI
Nebulizer Gas Flow: 1.5L/min
CDL: -20.0v
CDL Temp: 250°C
Block Temp: 200°C

Probe bias: +4.5kv
Detector: 1.5kv
T. Flow: 0.2ml/min
B. conc: 50% H₂O/50% ACN